



Designation: B748 – 90 (Reapproved 2021)

Standard Test Method for Measurement of Thickness of Metallic Coatings by Measurement of Cross Section with a Scanning Electron Microscope¹

This standard is issued under the fixed designation B748; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the measurement of metallic coating thicknesses by examination of a cross section with a scanning electron microscope (SEM).

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

E3 Guide for Preparation of Metallographic Specimens

E766 Practice for Calibrating the Magnification of a Scanning Electron Microscope

3. Summary of Test Method

3.1 A test specimen is cut, ground, and polished for metallographic examination by an SEM of a cross section of the coating. The measurement is made on a conventional micrograph or on a photograph of the video waveform signal for a single scan across the coating.

¹ This test method is under the jurisdiction of ASTM Committee B08 on Metallic and Inorganic Coatings and is the direct responsibility of Subcommittee B08.10 on Test Methods.

Current edition approved Oct. 1, 2021. Published November 2021. Originally approved in 1985. Last previous edition approved in 2016 as B748 – 90(2016). DOI: 10.1520/B0748-90R21.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

4. Significance and Use

4.1 This test method is useful for the direct measurement of the thicknesses of metallic coatings and of individual layers of composite coatings, particularly for layers thinner than normally measured with the light microscope.

4.2 This test method is suitable for acceptance testing.

4.3 This test method is for the measurement of the thickness of the coating over a very small area and not of the average or minimum thickness per se.

4.4 Accurate measurements by this test method generally require very careful sample preparation, especially at the greater magnifications.

4.5 The coating thickness is an important factor in the performance of a coating in service.

5. Equipment

5.1 The scanning electron microscope shall have a resolution of at least 50 nm. Suitable instruments are available commercially.

6. Factors Affecting the Measurement Reliability

6.1 *Surface Roughness*—If the coating or its substrate is rough relative to the coating thickness, one or both of the interfaces bounding the coating cross section may be too irregular to permit accurate measurement of the average thickness in the field of view.

6.2 *Taper of Cross Section*—If the plane of the cross section is not perpendicular to the plane of the coating, the measured thickness will be greater than the true thickness. For example, an inclination of 10° to the perpendicular will contribute a 1.5 % error. True thickness, (t), equals measured thickness, (t_m), multiplied by the cosine of the angle of inclination (θ): $t = t_m \times \cos(\theta)$. (See X1.3.2.)

6.3 *Specimen Tilt*—Any tilt of the specimen (plane of the cross section) with respect to the SEM beam, may result in an erroneous measurement. The instrument should always be set for zero tilt.

6.4 *Oblique Measurement*—If the thickness measurement is not perpendicular to the plane of the coating, even when there

is no taper (6.2) or tilt (6.3), the measured value will be greater than the true thickness. This consideration applies to the conventional micrograph (9.3.1) and to the direction of the single video waveform scans (9.3.2).

6.5 Deformation of Coating—Detrimental deformation of the coating can be caused by excessive temperature or pressure during the mounting and preparation of cross sections of soft coatings.

6.6 Rounding of Edge of Coating—If the edge of the coating cross section is rounded, that is, if the coating cross section is not completely flat up to its edges, the observed thickness may differ from the true thickness. Edge rounding can be caused by improper mounting, grinding, polishing, or etching.

6.7 Overplating of Specimen—Overplating of the test specimen serves to protect the coating edges during preparation of cross sections and thus to prevent an erroneous measurement. Removal of coating material during surface preparation for overplating can cause a low thickness measurement.

6.8 Etching—Optimum etching will produce a clearly defined and narrow dark line at the interface of two metals. A wide or poorly defined line can result in an inaccurate measurement.

6.9 Smearing—Polishing may leave smeared metal that obscures the true boundary between the two metals and results in an inaccurate measurement. This may occur with soft metals like lead, indium, and gold. To help identify whether or not there is smearing, repeat the polishing, etching, and measurement several times. Any significant variations in readings indicates possible smearing.

6.10 Poor Contrast—The visual contrast between metals in the SEM is poor when their atomic numbers are close together. For example, bright and semibright nickel layers may not be discriminable unless their common boundary can be brought out sufficiently by appropriate etching and SEM techniques. For some metal combinations, energy dispersive X-ray techniques (see X1.4.5) or backscatter image techniques (see X1.4.6) may be helpful.

6.11 Magnification:

6.11.1 For any given coating thickness, measurement errors tend to increase with decreasing magnification. If practical, the magnification should be chosen so that the field of view is between 1.5 and 3× the coating thickness.

6.11.2 The magnification readout of an SEM is often poorer than the 5 % accuracy often quoted and the magnification has been found for some instruments to vary by 25 % across the field. Magnification errors are minimized by appropriate use of an SEM stage micrometer and appropriate experimental procedure (see Practice E766).

6.12 Uniformity of Magnification—Because the magnification may not be uniform over the entire field, errors can occur if both the calibration and the measurement are not made over the same portion of the field. This can be very important.

6.13 Stability of Magnification:

6.13.1 The magnification of an SEM often changes or drifts with time. This effect is minimized by mounting the stage

micrometer and test specimen side by side on the SEM stage so as to keep the transfer time short.

6.13.2 A change in magnification can occur when adjustments are made with the focusing and other electronic SEM controls. Such a change is prevented by not using the electronic focus controls or other electronic SEM controls after photographing the stage micrometer scale except to focus with the mechanical X, Y, and Z controls. Appropriate manipulation of the X, Y, and Z controls will bring the specimen surface to the focal point of the SEM beam.

6.14 Stability of Micrographs—Dimensional changes of micrographs can take place with time and with temperature and humidity changes. If the calibration micrograph of the stage micrometer scale and the micrograph of the test specimen are kept together and time is allowed for stabilization of the photographic paper, errors from this source will be minimized.

7. Preparations of Cross Sections

7.1 Prepare, mount, grind, polish, and etch the test specimen so that the following occurs:

7.1.1 The cross section is perpendicular to the plane of the coating,

7.1.2 The surface is flat and the entire width of the coating image is simultaneously in focus at the magnification to be used for the measurement,

7.1.3 All material deformed by cutting or cross sectioning is removed,

7.1.4 The boundaries of the coating cross section are sharply defined by contrasting appearance, or by a narrow, well-defined line, and

7.1.5 If the video waveform signal is to be measured, the signal trace is flat except across the two boundaries of the coating.

7.2 For further guidance see [Appendix X1](#).

8. Calibration of Magnification

8.1 Calibrate the SEM with an SEM stage micrometer and determine the magnification factor, M , in accordance with Practice E766 (see X1.4.2). Other calibration methods may be used if it can be demonstrated that they are sufficiently accurate for meeting the requirement of Section 12.

8.2 If practical, the stage micrometer and the test specimen shall be mounted side by side on the SEM stage.

9. Procedure

9.1 Operate the SEM in accordance with the manufacturer's instructions.

9.2 Take into account the factors listed in Sections 6 and 12.

9.3 Make a micrograph of the test specimen under the same conditions and instrument settings as used for the calibration and make an appropriate measurement of the micrograph image. Carry out this step in accordance with 9.3.1 or 9.3.2.

9.3.1 Conventional Micrograph:

9.3.1.1 With the boundaries of the coating clearly and sharply defined, make conventional micrographs of the SEM stage micrometer scale and of the test specimen.