



Designation: E3161 – 21

Standard Practice for Preparing a *Pseudomonas aeruginosa* or *Staphylococcus aureus* Biofilm using the CDC Biofilm Reactor¹

This standard is issued under the fixed designation E3161; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This practice specifies the parameters for growing a *Pseudomonas aeruginosa* (ATCC 15442) or *Staphylococcus aureus* (ATCC 6538) biofilm that can be used for disinfectant efficacy testing using the Test Method for Evaluating Disinfectant Efficacy Against *Pseudomonas aeruginosa* Biofilm Grown in CDC Biofilm Reactor Using Single Tube Method (E2871) or in an alternate method capable of accommodating the coupons used in the CDC Biofilm Reactor. The resulting biofilm is representative of generalized situations where biofilm exist on hard, non-porous surfaces under shear rather than being representative of one particular environment. Additional bacteria may be grown using the basic procedure outlined in this document, however, alternative preparation procedures for frozen stock cultures and biofilm generation (for example, medium concentrations, baffle speed, temperature, incubation times, coupon types, etc.) may be necessary.

1.2 This practice uses the CDC Biofilm Reactor created by the Centers for Disease Control and Prevention (1).² The CDC Biofilm Reactor is a continuously stirred tank reactor (CSTR) with high wall shear. The reactor is versatile and may also be used for growing or characterizing various species of biofilm, or both (2-4) provided appropriate adjustments are made to the growth media and operational parameters of the reactor.

1.3 Basic microbiology training is required to perform this practice.

1.4 *Units*—The values stated in SI units are to be regarded as standard. No other units of measurement are included in this practice.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate*

appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*³

E2756 Terminology Relating to Antimicrobial and Antiviral Agents

E2871 Test Method for Determining Disinfectant Efficacy Against Biofilm Grown in the CDC Biofilm Reactor Using the Single Tube Method

3. Terminology

3.1 *Definitions:*

3.1.1 For definition of terms used in this method refer to Terminology E2756.

3.1.2 *batch phase, n*—establishment of the biofilm by operating the reactor without the flow of nutrients (batch phase growth medium), but with mixing.

3.1.3 *biofilm, n*—microorganisms living in a self-organized community attached to surfaces, interfaces, or each other, embedded in a matrix of extracellular polymeric substances of microbial origin, while exhibiting altered phenotypes with respect to growth rate and gene transcription.

3.1.4 *continuously stirred tank reactor (CSTR) phase, n*—establishment of a steady state biofilm population achieved with the continuous flow of nutrients (continuous flow growth medium) in a glass vessel.

3.1.5 *coupon, n*—biofilm sample surface.

4. Summary of Practice

4.1 This practice is used for growing a *P. aeruginosa* or *S. aureus* biofilm in the CDC Biofilm Reactor. The biofilm is

¹ This practice is under the jurisdiction of ASTM Committee E35 on Pesticides, Antimicrobials, and Alternative Control Agents and is the direct responsibility of Subcommittee E35.15 on Antimicrobial Agents.

Current edition approved Nov. 1, 2021. Published January 2022. Originally approved in 2018. Last previous edition approved in 2018 as E3161–18. DOI: 10.1520/E3161–21.

² The boldface numbers in parentheses refer to the list of references at the end of this standard.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

established by operating the reactor in batch phase (no flow of the nutrients) for 24 h. A steady state population is reached while the reactor operates for an additional 24 h with continuous flow of the nutrients. The residence time of the nutrients in the reactor is set to select for biofilm growth, and is species and reactor parameter specific. During the entire 48 h, the biofilm is exposed to continuous fluid shear from the rotation of a baffled stir bar. Controlling the rate at which the baffle turns determines the intensity of the shear stress to which the coupons are exposed. At the end of the 48 h, the biofilm-laden coupons are used for testing.

5. Significance and Use

5.1 Bacteria that exist in biofilms are phenotypically different from suspended cells of the same genotype. Research has shown that biofilm bacteria are more difficult to kill than suspended bacteria (4, 5). Laboratory biofilms are engineered in growth reactors designed to produce a specific biofilm type. Altering system parameters will correspondingly result in a change in the biofilm. The purpose of this practice is to direct a user in the growth of a *P. aeruginosa* or *S. aureus* biofilm by clearly defining the operational parameters to grow a biofilm that can be assessed for efficacy using the Standard Test Method for Evaluating Disinfectant Efficacy Against *Pseudomonas aeruginosa* Biofilm Grown in CDC Biofilm Reactor Using Single Tube Method (E2871).

5.2 Operating the CDC Biofilm Reactor at the conditions specified in this method generates biofilm at log densities ($\log_{10}$ CFU per coupon) ranging from 8.0 to 9.5 for *P. aeruginosa* and 7.5 to 9.0 for *S. aureus*. These levels of biofilm are anticipated on surfaces conducive to biofilm formation such as the conditions outlined in this method.

5.2.1 To achieve an *S. aureus* biofilm with a population comparable to that for *P. aeruginosa* using the bacterial liquid growth medium conditions specified here, the *S. aureus* biofilm must be grown at $36\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ rather than at room temperature ($21\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$).

6. Apparatus

6.1 *Culture Tubes and Culture Tube Closures*—any glass or plastic tube with a volume capacity of at least 15 mL.

6.2 *Calibrated Pipetter*—continuously adjustable pipetter with volume capability of 1 mL.

6.3 *Vortex*—any vortex that will ensure proper agitation and mixing of culture tubes.

6.4 *Ultrasonic Water Bath*—any cavitating sonicating bath that operates at 45 ± 5 kHz and which has a volume large enough to accommodate 50 mL or 250 mL conical tubes.

6.5 *Analytical Balance*—sensitive to 0.01 g.

6.6 *Sterilizer*—any steam sterilizer that can produce the conditions of sterilization is acceptable.

6.7 *Peristaltic Pump*—pump head that can hold size 16 or equivalent peristaltic pump tubing. Use a separate pump for each reactor.

6.8 *Digital Magnetic Stir Plate*—top plate of at least 10.16 cm by 10.16 cm that can rotate at a range of 60 r/min to 125 r/min ± 5 r/min.

6.9 *Silicone Tubing*—multiple sizes: size 16 tubing or equivalent designed for use in a peristaltic pump (used for most connections between CSTR growth medium carboy and the reactor), and size 18 or 25 tubing or equivalent (used for reactor effluent). All sizes must withstand sterilization (for example, platinum cured).

6.10 *Norprene Tubing (or equivalent)*—size 16 or equivalent Norprene tubing. Recommended for use in the peristaltic pump.

6.11 *Glass Flow Break*—any that will connect with size 16 tubing and withstand sterilization, used to prevent microbial contamination of the nutrient reservoir from the biofilm reactor.

6.11.1 *Clamp*—used to hold flow break, extension clamp with 0.5 cm minimum grip size.

6.11.2 *Clamp Stand*—height no less than 76.2 cm, used with clamp to suspend glass flow break vertically and stabilize tubing above reactor.

6.12 *Reactor Components*.⁴

6.12.1 *Berzelius Borosilicate Glass Tall Beaker*—1000 mL without pour spout, 9.5 cm ± 0.5 cm diameter. Barbed outlet spout added at 400 mL ± 50 mL mark. Spout angled 30° to 45° to ensure drainage. Spout should accommodate size 18 or 25 flexible silicone tubing.

6.12.2 *Reactor Top*—Fig. 1. Ultra-high molecular weight (UHMW) polyethylene top (10.1 cm diameter tapering to 8.33 cm) equipped with a minimum of 3 holes accommodating 6 to 8 cm long pieces of stainless steel or other rigid autoclavable tubing with outside diameter of 5 mm to 8 mm for medium inlet, air exchange and inoculation port. Center hole, 1.27 cm diameter, to accommodate the glass rod used to support the baffle assembly. Eight rod holes, 1.905 cm diameter, notched to accommodate stainless steel rod alignment pin (0.236 cm outside diameter). O-ring, attached to underside of reactor top.

6.12.3 *Polypropylene Rods*—Fig. 2. Eight polypropylene rods, 21.08 cm long, two types: coupon holder machined to hold three coupons (see 6.12.4) at the immersed end, three 316 stainless steel set screws embedded in side to hold coupons in place; and coupon holder blanks, without coupon recesses. Rods fit into holes in reactor top and lock into preformed notches with alignment pin.

6.12.4 *Coupons*—twenty-four cylindrical coupons (for example, borosilicate glass) with a diameter of 1.27 cm ± 0.013 cm, thickness of approximately 3.0 mm.

6.12.5 *Small Allen Wrench (1.27 mm, hex)*—for adjusting set screws.

⁴ The sole source of supply of the apparatus (CDC Biofilm Reactor) and associated coupons known to the committee at this time is BioSurface Technologies, Corp. www.biosurfaces.biz. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend. The user may also build the reactor.