



Designation: G175 – 13 (Reapproved 2021)

Standard Test Method for Evaluating the Ignition Sensitivity and Fault Tolerance of Oxygen Pressure Regulators Used for Medical and Emergency Applications¹

This standard is issued under the fixed designation G175; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 For the purpose of this standard, a pressure regulator, also called a pressure-reducing valve, is a device intended for medical or emergency purposes that is used to convert a medical or emergency gas pressure from a high, variable pressure to a lower, more constant working pressure [21 CFR 868.2700 (a)]. Some of these oxygen pressure regulators are a combination of a pressure regulator and cylinder valve. These devices are often referred to as valve integrated pressure regulators, or VIPRs.

1.2 This standard provides an evaluation tool for determining the ignition sensitivity and fault tolerance of oxygen pressure regulators and VIPRs used for medical and emergency applications. An ignition-sensitive pressure regulator or VIPR is defined as having a high probability of ignition as evaluated by rapid pressurization testing (Phase 1). A fault-tolerant pressure regulator or VIPR is defined as having a low consequence of ignition as evaluated by forced ignition testing (Phase 2).

NOTE 1—It is essential that a risk assessment be carried out on breathing gas systems, especially concerning toxic product formation due to ignition or decomposition of nonmetallic materials as weighed against the risk of flammability (refer to Guide G63 and ISO 15001.2). See Appendix X1 and Appendix X2 for details.

1.3 This standard applies only to:

1.3.1 Oxygen pressure regulators used for medical and emergency applications that are designed and fitted with CGA 540 inlet connections, CGA 870 pin-index adapters (CGA V-1), or EN ISO 407 pin-index adapters.

1.3.2 Oxygen VIPRs used for medical and emergency applications that are designed to be permanently fitted to a medical gas cylinder.

1.4 This standard is a test standard not a design standard; *This test standard is not intended as a substitute for traditional*

design requirements for oxygen cylinder valves, pressure regulators and VIPRs. A well-designed pressure regulator or VIPR should consider the practices and materials in standards such as Guides G63, G88, G94, and G128, Practice G93, CGA E-18, CGA E-7, ISO 15001, ISO 10524-1 and ISO 10524-3.

NOTE 2—Medical applications include, but are not limited to, oxygen gas delivery in hospitals and home healthcare, and emergency applications including, but not limited to, oxygen gas delivery by emergency personnel.

1.5 This standard is also intended to aid those responsible for purchasing or using oxygen pressure regulators and VIPRs used for medical and emergency applications by ensuring that selected pressure regulators are tolerant of the ignition mechanisms that are normally active in oxygen systems.

1.6 This standard does not purport to address the ignition sensitivity and fault tolerance of an oxygen regulator or VIPR caused by contamination during field maintenance or use. Pressure regulator and VIPR designers and manufacturers should provide design safeguards to minimize the potential for contamination or its consequences (see Guide G88).

NOTE 3—Experience has shown that the use of bi-direction flow filters in components can lead to accumulation and re-release of contaminants (refer to Guide G88-05 Section 7.5.3.8 and EIGA Info 21/08).

1.7 The values stated in inch-pound units are to be regarded as standard. The values given in parentheses are mathematical conversions to SI units that are provided for information only and are not considered standard.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee G04 on Compatibility and Sensitivity of Materials in Oxygen Enriched Atmospheres and is the direct responsibility of Subcommittee G04.01 on Test Methods.

Current edition approved Dec. 1, 2021. Published December 2021. Originally published as PS 127 – 00. Last published in 2013 as G175 – 13. DOI: 10.1520/G0175-13R21.

2. Referenced Documents

2.1 ASTM Standards:²

- G63 Guide for Evaluating Nonmetallic Materials for Oxygen Service
- G88 Guide for Designing Systems for Oxygen Service
- G93 Guide for Cleanliness Levels and Cleaning Methods for Materials and Equipment Used in Oxygen-Enriched Environments
- G94 Guide for Evaluating Metals for Oxygen Service
- G128 Guide for Control of Hazards and Risks in Oxygen Enriched Systems
- D618 Practice for Conditioning Plastics for Testing
- D4066 Classification System for Nylon Injection and Extrusion Materials (PA)
- D6779 Classification System for and Basis of Specification for Polyamide Molding and Extrusion Materials (PA)

2.2 Other ASTM Documents:²

- Manual 36 Safe Use of Oxygen and Oxygen Systems
- Smith, S. R., and Stoltzfus, J. M., "Preliminary Results of ASTM G175 Interlaboratory Studies," *Flammability and Sensitivity of Materials in Oxygen-Enriched Atmospheres: Tenth Volume, ASTM STP 1454*, T. A. Steinberg, H. D. Beeson, and B. E. Newton, Eds., ASTM International, West Conshohocken, PA, 2003.
- Smith, S. R., and Stoltzfus, J. M., "ASTM G175 Interlaboratory Study on Forced Ignition Testing," *Journal of ASTM International*, Vol. 3, No. 7, Paper ID JAI13542, pp. 314-318.

2.3 Compressed Gas Association (CGA) Standards:³

- CGA E-4 Standard for Gas Pressure Regulators
- CGA E-7 Standard for Medical Pressure Regulators
- CGA E-18 Medical Gas Valve Integrated Pressure Regulators
- CGA G-4 Oxygen
- CGA G-4.1 Cleaning Equipment for Oxygen Service
- CGA V-1 American National/Compressed Gas Association Standard for Compressed Gas Cylinder Valve Outlet and Inlet Connections
- CGA V-14 Performance Standard for Sealing Gaskets Used on CGA 870 Connections for Medical Oxygen Service

2.4 United States Pharmacopeial Convention Standard:⁴

- USP 24 – NF 19 Oxygen monograph

2.5 Federal Regulation:⁵

- 21 CFR 868.2700 (a) Pressure regulator

2.6 ISO Standards:⁶

- ISO 10524-1 Pressure regulators for use with medical gases — Part 1: Pressure regulators and pressure regulators with flow-metering devices
- ISO 10524-3 Pressure regulators for use with medical gases — Part 3: Pressure regulators integrated with cylinder valves
- ISO 15001 Anaesthetic and respiratory equipment – Compatibility with oxygen

2.7 European Industrial Gas Association Documents:⁷

- EIGA Info 21/08 Cylinder Valves—Design Considerations

3. Summary of Test Method

3.1 This test method comprises two phases. A pressure regulator or VIPR must pass both phases in order to be considered ignition-resistant and fault-tolerant.

3.2 *Phase 1: Oxygen Pressure Shock Test*—In this test phase, the ignition sensitivity of the pressure regulator design is evaluated by subjecting the pressure regulator or VIPR to heat from oxygen pressure shocks. The test is performed according to ISO 10524-1 Section 6.6 for oxygen regulators, which is similar to CGA E-7 and ISO 10524-3 Section 6.6 for oxygen VIPRs.

3.3 *Phase 2: Promoted Ignition Test*—The Phase 1 component test system is used for Phase 2 to pressure shock a pressure regulator or VIPR so that an ignition pill is kindled to initiate combustion within the pressure regulator or VIPR. The ignition source is representative of severe, but realistic, service conditions.

3.3.1 *Oxygen Pressure Regulator*—In this test phase, and for this component type, fault tolerance is evaluated by subjecting the pressure regulator to the forced application of a positive ignition source at the pressure regulator inlet to simulate cylinder valve seat ignition and particle impact events.

3.3.2 *Oxygen VIPR*—In this test phase and for this component type, fault tolerance is evaluated by subjecting the VIPR to the forced application of a positive ignition source at the cylinder connection port to simulate a shut-off valve seat ignition and particle impact events in the use (not cylinder filling mode) configuration.

4. Significance and Use

4.1 This test method comprises two phases and is used to evaluate the ignition sensitivity and fault tolerance of oxygen pressure regulators used for medical and emergency applications.

4.2 *Phase 1: Oxygen Pressure Shock Test*—The objective of this test phase is to determine whether the heat or temperature from oxygen pressure shocks will result in burnout or visible heat damage to the internal parts of the pressure regulator.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from Compressed Gas Association (CGA), 4221 Walney Rd., 5th Floor, Chantilly, VA 20151-2923, <http://www.cganet.com>.

⁴ Available from U.S. Pharmacopeia (USP), 12601 Twinbrook Pkwy., Rockville, MD 20852.

⁵ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.

⁶ Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, Case postale 56, CH-1211, Geneva 20, Switzerland, <http://www.iso.ch>.

⁷ Available from European Industrial Gas Association (EIGA), AISBL Avenue des Arts, 3-5-b-1210 Brussels, Belgium, <https://www.eiga.eu/>.