



Published 1980-04-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Glass alcoholometers and alcohol hydrometers not incorporating a thermometer

ERRATUM

Page 1

Replace table 1 by the following :

TABLE 1 — Classes of accuracy

Class	Minimum mean distance between centres of adjacent graduation lines mm	Type
1	1,5	1, 2
2	1,06	1, 2, 3



Glass alcoholometers and alcohol hydrometers not incorporating a thermometer

Alcomètres et aréomètres pour alcool, sans thermomètre incorporé

First edition — 1979-06-01

UDC 542.3 : 531.756.3

Ref. No. ISO 4801-1979 (E)

Descriptors : laboratory glassware, alcoholometers, hydrometers, specifications.